

## **NeoBridge**<sup>®</sup>

**Umbilical Catheter Holder** 



## FEATURES AND BENEFITS

- Skin friendly NeoBond® Hydrocolloid base
- Secure positioning flaps designed to relieve stress on tubes and help prevent dislodgement
- Accommodates two catheters
- Flap designed to be opened and closed repeatedly
- Clear flap allows for better visibility
- Not made with natural rubber latex or plasticizer DEHP
- Individually packaged

Cat No.	Size	Approx Wt	Qty/Unit
N700	Micro	800 or less	20/box
N701	Small	800-1200 gm	20/box
N702	Large	1200-2200 gm	20/box



U.S. Patent #5,370,627

## NeoBridge® Umbilical Catheter Holder

## **INDICATIONS FOR USE**

The NeoBridge is intended to anchor and support umbilical catheters. It is intended for use on pediatric (neonates, infants and children) patients.

## **DIRECTIONS FOR USE**

## **Preparation**

After catheter is has been inserted, clean and dry skin per hospital protocol.

Note: It is important for skin to be as clean and dry as possible. NeoBridge will not adhere properly to moist skin or hair. Oils and lotions will also adversely affect adhesion.

DO NOT USE ALCOHOL.

## Application

## Step 1

Before removing liners, warm NeoBridge in hands for 60 seconds. May also be held under radiant or heel warmer for 10 to 15 seconds.

## Step 2

Discard white, paper liner, then peel and discard clear, plastic liner from hydrocolloid base.

## Step 3

With slit on right side and wider end of base facing patient's head, slip NeoBridge around catheter.

## Step 4

Apply base to skin and hold in place for 60 seconds to ensure proper adhesion.

## Step 5

Peel and discard frosted liner from top flap. Bring flap up and adhere to catheter.

Note: Catheter should exit side of flap. Do not overlap green tab. If using two catheters, have one catheter exit each side of flap.

## Step 6

Peel and discard frosted liner from bottom flap. Bring flap up to align with top flap and press firmly together, enclosing catheter(s) between adhesives.

## Step 7

Pull flaps apart to remove or reposition catheter(s) as needed, then close to secure.

## Removal

## Step 1

Replace NeoBridge every 5 days or per hospital protocol, whichever is sooner.

## Step 2

Pull flaps apart to remove catheter(s).

## Step 3

Saturate NeoBridge with water or saline.

#### Step 4

Slowly peel NeoBridge away from skin as you swab with water or saline.

## **GENERAL INFORMATION**

## **Product Materials**

NeoBond Hydrocolloid, double coated adhesive (does not contact skin), polyurethane film Made with natural rubber latex: No Made with plasticizer DEHP: No Consult MSDS for complete material information.

## Size

N700 Micro (Approximate weight: 800 grams or less) N701 Small (Approximate weight: 800 to 1200 grams) N702 Large (Approximate weight: 1200 to 2200 grams)

## Packaging

Individually packaged: Yes Package type: Polyethylene bag Standard box, no dispense feature

Visit neotechproducts.com for the most up-to-date information or to download the Material Safety Data Sheet (MSDS).

Polyethylene bags are biodegradable. Boxes are printed with environmentally friendly soy ink on recyclable material.

## **CAUTIONS & WARNINGS**

Federal law restricts this device to sale by or on the order of a physician.

NeoBridge should only be used while the patient is under the continuous, direct supervision of trained healthcare professionals.

Discontinue immediately if skin irritation occurs.

NeoBridge will not adhere properly to moist skin or hair. Oils and lotions will adversely affect adhesion.

Do not use alcohol to clean skin prior to application. Single patient use only.

# Made in USA

# For more information or to order your FREE sample call **1-800-966-0500** or visit **neotechproducts.com**

Neotech Products, Inc. • 28430 Witherspoon Pkwy • Valencia, CA 91355, USA